




CERTIFICATE OF INSPECTION

Evaluation Profile



Work Name	Sample	Operator	Mitutoyo
Measuring Tool	SurfTest SJ-210	Comment	Ver1.00
Standard	ISO 1997	N	5
Profile	R	Cut-Off	2.5mm
λs	8.0um	Filter	GAUSS
Ra	2.958 um		